



# Memo

**To:** Evelyn Foust  
Communicable Disease Branch Head  
Cordelia Williams

**From:** Laboratory Improvement Coordinator  
Pete Moore  
Ron Higginbotham  
NCSLPH Lab Improvement Staff

**CC:** Local Health Department Lab Managers

**Date:** April 14, 2025

**Re:** GC Lot Status

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Modified Thayer Martin II: **Lot Number: 653386** Expiration Date : **05/23/25**  
Along with **Bio-Bag™ Type C**: Lot Number: **B02G058M** Expiration Date: **06/30/25** and  
Lot Number: **B02G100M** Expiration Date: **11/30/25**

**Summary of NCSLPH QC results for this Lot #:**

<i>N. gonorrhoeae</i> ATCC# 43069 -	<i>S. epidermidis</i> ATCC# 12228 – No Growth
	<i>P. mirabilis</i> ATCC# 43071 – No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

**Lot number failed:** None  
Hardy Diagnostics Modified Thayer Martin

\*\*If your contact information changes, please provide that information to Tracey Shives in Laboratory Improvement at (336)-306-4302 or via email at [tracey.shives@dhhs.nc.gov](mailto:tracey.shives@dhhs.nc.gov). Thank you.

NC DEPARTMENT OF HEALTH AND HUMAN SERVICES • DIVISION OF PUBLIC HEALTH • STATE LABORATORY OF PUBLIC HEALTH

LOCATION: 4312 District Drive, Raleigh, NC 27607

MAILING ADDRESS: 1918 Mail Service Center, Raleigh, NC 27699-1918

www.ncdhhs.gov • <http://publichealth.nc.gov> • <https://slph.dph.ncdhhs.gov/> • TEL: 919-733-7834 • FAX: 919-733-8695

AN EQUAL OPPORTUNITY / AFFIRMATIVE ACTION EMPLOYER