



NC DEPARTMENT OF
**HEALTH AND
HUMAN SERVICES**
Division of Public Health

JOSH STEIN • Governor

DEVPUTTA SANGVAI • Secretary

KAREN BURKES • Acting Deputy Secretary for Health

KELLY KIMPLE • Acting Director, Division of Public Health

SCOTT M. SHONE, PhD, HCLD (ABB) • Laboratory Director, State
Laboratory of Public Health

Memo

To: Evelyn Foust
Communicable Disease Branch Head
Cordelia Williams
From: Laboratory Improvement Coordinator
Pete Moore
Ron Higginbotham
NCSLPH Lab Improvement Staff
CC: Local Health Department Lab Managers
Date: May 14, 2025
Re: GC Lot Status

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Modified Thayer Martin II: **Lot Number: 656986** Expiration Date: **07-03-2025**
Along with **Bio-Bag™ Type C**: Lot Number: **B02G100M** Expiration Date: **11-30-2025** and
Lot Number: **B02G058M** Expiration Date: **06-30-2025**

Summary of NCSLPH QC results for this Lot #:

<i>N. gonorrhoeae</i> ATCC# 43069 -	<i>S. epidermidis</i> ATCC# 12228 – No Growth
	<i>P. mirabilis</i> ATCC# 43071 – No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

Lot number failed: None
Hardy Diagnostics Modified Thayer Martin

****If your contact information changes, please provide that information to Tracey Shives in Laboratory Improvement at (336)-306-4302 or via email at tracey.shives@dhhs.nc.gov. Thank you.**

NC DEPARTMENT OF HEALTH AND HUMAN SERVICES • DIVISION OF PUBLIC HEALTH • STATE LABORATORY OF PUBLIC HEALTH

LOCATION: 4312 District Drive, Raleigh, NC 27607

MAILING ADDRESS: 1918 Mail Service Center, Raleigh, NC 27699-1918

www.ncdhhs.gov • <http://publichealth.nc.gov> • <https://slph.dph.ncdhhs.gov/> • TEL: 919-733-7834 • FAX: 919-733-8695

AN EQUAL OPPORTUNITY / AFFIRMATIVE ACTION EMPLOYER