



# Memo

**To:** Evelyn Foust  
Communicable Disease Branch Head  
**From:** Cordelia Williams  
Laboratory Improvement Coordinator  
Pete Moore  
Ron Higginbotham  
NCSLPH Lab Improvement Staff  
**CC:** Local Health Department Lab Managers  
**Date:** January 12, 2026  
**Re:** GC Lot Status

---

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are accepted for use:

Modified Thayer Martin II: **Lot Number: 675814** Expiration Date: **02-27-26**  
Along with **Bio-Bag™ Type C**: Lot Number: **B04G311M** Expiration Date: **03-31-26** and  
Lot Number: **B03G340M** Expiration Date: **03-31-26**

### Summary of NCSLPH QC results for this Lot #:

<i>N. gonorrhoeae</i> ATCC# 43069 – Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
	<i>P. mirabilis</i> ATCC# 43071 – No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

**Lot number failed:** None  
Hardy Diagnostics Modified Thayer Martin

\*\*If your contact information changes, please provide that information to Tracey Shives in Laboratory Improvement at (336)-306-4302 or via email at [tracey.shives@dhhs.nc.gov](mailto:tracey.shives@dhhs.nc.gov). Thank you.

NC DEPARTMENT OF HEALTH AND HUMAN SERVICES • DIVISION OF PUBLIC HEALTH • STATE LABORATORY OF PUBLIC HEALTH

LOCATION: 4312 District Drive, Raleigh, NC 27607

MAILING ADDRESS: 1918 Mail Service Center, Raleigh, NC 27699-1918

[www.ncdhhs.gov](http://ncdhhs.gov) • <http://publichealth.nc.gov> • <https://slph.dph.ncdhhs.gov/> • TEL: 919-733-7834 • FAX: 919-733-8695

AN EQUAL OPPORTUNITY / AFFIRMATIVE ACTION EMPLOYER