



NC DEPARTMENT OF
**HEALTH AND
HUMAN SERVICES**
Division of Public Health

JOSH STEIN • Governor

DEVPUTTA SANGVAI • Secretary

DEBRA FARRINGTON • Deputy Secretary for Health

KELLY KIMPLE • Director, Division of Public Health

SCOTT M. SHONE, PhD, HCLD (ABB) • Laboratory Director, State
Laboratory of Public Health

GC Media Evaluation Results

To: North Carolina Health Department Laboratory Managers

From: Laboratory Improvement, NC State Laboratory of Public Health

Date: July 10, 2026

Re: The Quality Control (QC) of Modified Thayer Martin (MTM) II Media

Quality Control Evaluation of GC Media

According to the NC State Laboratory of Public Health (SLPH) quality control evaluation, the following lots of media and bio-bags were found acceptable for use of presumptive gonorrhoeae culture testing.

Lots approved:

- **Modified Thayer Martin (MTM) II Media**
 - Lot: **688215P** Expiration Date: **08/25/2026**
- **BD Type C Bio-bag (for transport use only)**
 - Lot: **B01H282M** Expiration Date: **12/31/2026**

Summary of the NC SLPH QC Evaluation Results

The following quality control test were conducted; the results of these tests are listed below.

- Sterility Check: **Acceptable**
- Visual Check: **Acceptable**
- Nutritive Test:
 - Neisseria gonorrhoeae ATCC® 43069: **Growth**
- Inhibition Test:
 - Staphylococcus epidermidis ATCC®12228: **Complete inhibition**
 - Proteus mirabilis ATCC® 43071: **Complete inhibition**

Questions and Concerns

Please direct all questions regarding QC, media and Bio-bags to the Laboratory Improvement Coordinator:

- Cordelia Williams, email: Cordelia.williams@dhhs.nc.gov, phone: (919)807-8816

NC DEPARTMENT OF HEALTH AND HUMAN SERVICES • DIVISION OF PUBLIC HEALTH • STATE LABORATORY OF PUBLIC HEALTH

LOCATION: 4312 District Drive, Raleigh, NC 27607

MAILING ADDRESS: 1918 Mail Service Center, Raleigh, NC 27699-1918

www.ncdhhs.gov • <http://publichealth.nc.gov> • <https://slph.dph.ncdhhs.gov/> • TEL: 919-733-7834 • FAX: 919-733-8695

AN EQUAL OPPORTUNITY / AFFIRMATIVE ACTION EMPLOYER